

Suggested kits	Solid	Porous	Liquids (open)	Liquids (hermetically sealed)	Deco Filter	Deco Pathogen	Medical EN285	Food
<b>Steam generator</b> <i>Generatore di vapore</i>	✓	✓	✓	✓	✓	✓	✓	✓
<b>Vacuum pump</b> <i>Pompa del vuoto</i>	▲	✓	○	○	✓	○	✓	✓
<b>Internal air compressor</b> <i>Compressore aria</i>	✓	✓	X	X	✓	X	✓	X
<b>Condensate cooling unit</b> <i>Raffreddatore condense sullo scarico</i>	✓	✓	✓	○	✓	✓	✓	✓
<b>Fast cooling unit</b> <i>Raffreddamento rapido</i>	○	○	✓	✓	○	✓	○	▲
<b>Chamber pressure tightness test</b> <i>Test tenuta camera</i>	○	○	○	○	○	✓	○	○
<b>Air -steam conuterpressure cycle</b> <i>Ciclo in contropressione aria-vapore</i>	○	○	○	✓	X	○	○	▲
<b>Safety thermal locking device</b> <i>Sistema interblocco termico</i>	○	○	▲	✓	○	▲	○	▲
<b>Decontamination cycle with discharge air filter</b> <i>Ciclo di decontaminazione con filtro scarico aria</i>	○	○	○	X	✓	X	○	○
<b>Decontamination cycle</b> <i>Ciclo di decontaminazione</i>	○	○	○	○	X	✓	○	○
<b>F<sub>0</sub> sterilization</b> <i>Sterilizzazione F<sub>0</sub></i>	○	○	○	○	○	○	○	✓
<b>EN 285 conformity</b> (vacuum pump, sterile air filter and external recorder) <i>Conformità EN 285 (pompa del vuoto, filtro aria sterile e registratore esterno)</i>	○	○	○	○	○	○	✓	○
<b>Additional PT100 probe</b> <i>Sonda PT100 aggiuntiva</i>	○	○	▲	▲	▲	▲	○	▲
<b>Process printer</b> <i>Stampante di processo</i>	○	○	○	○	○	○	✓	○
<b>Slow vacuum</b> <i>Vuoto lento</i>	○	○	○	○	○	○	○	✓
<b>Standard packing</b> <i>Imballo standard</i>	✓	✓	✓	✓	✓	✓	✓	✓

X - not available/non disponibile

✓- included/incluso

○ - optional/opzionale

▲ - recommended/raccomandati