

## OPEN CELL TOP WITH CAP

The Open Cell Top with Cap is easily adaptable to the Hanson VDC. The cap can be removed between sampling time points to add or remove material. It is best used for less viscous materials because the cell top can be assembled prior to applying the dose. The larger donor chamber also allows more donor material to be used.



### OPEN CELL TOP WITH CAP COMPONENTS

